



2019

ANALYTICAL SERVICES

Effective Date: January 1, 2019

2773 Downhill Drive
Steamboat Springs, CO 80487
www.acz.com

Phone: (970) 879-6590
Toll Free: (800) 334-5493
Fax: (815) 301-3857
Email: sales@acz.com



Table of Contents

Category	Page No.
<i>Water and Wastewater</i>	
Wet Chemistry	3-4
Anions by Ion Chromatography	5
Physical Properties	5
Sample Preparation	6
Metals Analysis	6
Metals Detection Limits	7
Organic Analysis: GC, GC/MS, and HPLC	8
Radiochemistry: Isotopic and Metals Analysis by ICP-MS	9
<i>Soil and Overburden</i>	
Sample Preparation	10
Extractions	10
Digestions	11
Chemical Properties	11-12
Physical Properties	13
Metals Analysis	13
Metals Detection Limits	14
Organic Analysis: GC, GC/MS, and HPLC	15
Radiochemistry: Isotopic and Metals Analysis by ICP-MS	16
<i>Air Quality Analysis</i>	
Analysis: GC/MS and GC/TCD	17
<i>Biological Tissue Analysis</i>	
Sample Preparation	18
Analysis	18
<i>Solid and Hazardous Waste - RCRA</i>	
Hazardous Waste Characteristics	19
Underground Storage Tank (UST) Analysis	20
Appendix IX Constituents	20
<i>Regulatory Parameters and Packages</i>	
California Administrative Code (CAC)	21
Clean Water Act (CWA)	21
Colorado Department of Public Health and Environment (CDPHE)	21
Nevada Division of Environmental Protection (NDEP)	22-23
Safe Drinking Water Act (SDWA)	24-25
<i>Administrative Services</i>	
Data Reporting Formats	26
Data Reporting Charges	26
Electronic Data Deliverables	27
Labor	27
Additional Services: Rush Turnaround, Sample Storage and Small Sampling Event	27
<i>General Terms and Conditions</i>	28-31



**Water and Wastewater
Inorganic
Wet Chemistry**

Parameter	Method ¹	MDL (mg/L)	Price/Sample
Acidity	SM 2310 B Titrimetric	10	
Alkalinity	SM 2320 B Titrimetric		
Includes HCO ₃ , CO ₃ , OH and Total Alkalinity		2	
Biological Activity Reaction Test (BART)			
Iron Related Bacteria	HACH IRB-BART	0, 1	
Slime Forming Bacteria	HACH SLYM-BART	0, 1	
Sulfate Reducing Bacteria	HACH SRB-BART	0, 1	
Biological Oxygen Demand (BOD)	SM 5210B Electrode 5-day (20C)	2	
Carbonaceous (CBOD)	SM 5210B Electrode 5-day (20C)	2	
Set-Up Charge for less than 4 samples			
Bromide	EPA M300.0 Ion Chromatography	0.05	
Carbon			
Dissolved Inorganic (DIC)	SM 5310 B, Combustion/IR Detection	1	
Dissolved Organic (DOC)	SM 5310 B, Combustion/IR Detection	1	
Total Inorganic (TIC)	SM 5310 B, Combustion/IR Detection	1	
Total Organic (TOC)	SM 5310 B, Combustion/IR Detection	1	
Cation-Anion Balance (CAB)	Calculation		
Requires Alkalinity, Cl, SO ₄ , Ca, Mg, Na and K			
Chemical Oxygen Demand (COD)	EPA M410.4 Colorimetric	10	
Chloride	SM 4500 Cl-E Automated Colorimetric	0.5	
Chloride	EPA M300.0 Ion Chromatography	0.5	
Chromium			
Hexavalent	SM 3500 Cr- B/D Colorimetric	0.005	
Trivalent	Calculation		
Requires Total and Hexavalent		0.01	
Coliform Bacteria (Fecal)	SM 9222 D	1 (colony)/100mL	
Corrosivity (Langlier Index)	SM 2330 B Calculation		
Requires Alkalinity, Ca, TDS, Field pH and Field Temperature		0.01 SI unit	
Cyanide			
Free (MDL may vary with matrix)	D 6888-09/OIA-1677-09	5 - 10 ug/L	
Available (MDL may vary with matrix)	D 6888-04/OIA-1677-09	5 - 10 ug/L	
Thiocyanate (SCN)	SM 4500 CN-M	0.1	
Total, UV Digestion	D 7511-09	0.003	
Total	EPA M335.4 Automated Colorimetric	0.003	
Weak Acid Dissociable (WAD)	SM 4500 CN-I	0.003	
Fluoride	SM 4500 F-C	0.05	
Fluoride	EPA M300.0 Ion Chromatography	0.05	

1. The methods listed throughout this fee schedule are ACZ's standard methods. Most parameters can be analyzed using alternative methods, such as SM or AOAC. Call an ACZ representative for availability and pricing of alternative methods.



**Water and Wastewater
Inorganic
Wet Chemistry**

Parameter	Method	MDL (mg/L)	Price/Sample
Iron			
Ferrous (II)	SM 3500 Fe B	0.01	
Ferric (III)	Calculation		
	Requires Dissolved and Ferrous	0.01	
Langlier Index (see Corrosivity)			
Nitrogen			
Ammonia (as N), NH ₃	EPA M350.1 Automated Colorimetric	0.05	
Ammonia (as N), Un-Ionized	Calculation (requires NH ₃ , Field pH and Temp)	0.1	
Nitrate/Nitrite (as N), NO ₃ /NO ₂	EPA M353.2 Automated Colorimetric	0.02	
Nitrite (as N), NO ₂	EPA M353.2 Automated Colorimetric	0.01	
Nitrate (as N), NO ₃	Calculation (requires NO ₃ /NO ₂ and NO ₂)	0.01	
Organic	Calculation (requires TKN and NH ₃)	0.1	
Inorganic	Calculation (requires NO ₂ /NO ₃ and NH ₃)	0.02	
Total	Calculation (requires NO ₂ /NO ₃ and TKN)	0.1	
Total Kjeldahl (TKN)	EPA M351.2 Block Digest./Auto Color.	0.1	
Oil & Grease, Total Recoverable	EPA 1664A Hexane Extraction	2	
Phenols			
Dissolved (with distillation)	EPA M420.4 Automated Colorimetric	0.003	
Total (with distillation)	EPA M420.4 Automated Colorimetric	0.003	
Phosphorus			
Orthophosphate (as P)	EPA M365.1 Automated Colorimetric	0.02	
Total (as P) (with Persulfate digestion)	EPA M365.1 Automated Colorimetric	0.02	
Sulfate (SO ₄)	D 516-02/07, Turbimetric	1	
Sulfate (SO ₄)	EPA M300.0 Ion Chromatography	0.5	
Sulfate (SO ₄)	SM 4500 SO4-D, Gravimetric	10	
Sulfide (as H ₂ S)	SM4500 S2-D Colorimetric	0.02	
Sulfite (SO ₃)	EPA M377.1 Titrimetric	2	

**Water and Wastewater
Inorganic***Anions by Ion Chromatography*

Parameter	Method	MDL (mg/L)	Price/Sample
Bromide (Br)	EPA M300.0	0.05	
Chloride (Cl)	EPA M300.0	0.5	
Fluoride (F)	EPA M300.0	0.05	
Sulfate (SO ₄)	EPA M300.0	0.5	

Physical Properties

Parameter	Method	MDL (mg/L)	Price/Sample
Color	HACH Color Wheel	5 Co/Pt	
Conductance (EC)	SM 2510 B (Meter)	1 (umhos/cm)	
Hardness (requires Ca and Mg)	SM 2340 B Calculation	1.5	
pH	SM 4500 H+B (Meter)	0.1 (units)	
Residue			
Settleable (SS)	SM 2540 F Imhoff Cone	0.1 (mL/L/hr)	
Total (TS)	SM 2540 B Gravimetric, 105° C	10	
Filterable (TDS)	SM 2540 C Gravimetric, 180° C	10	
Non-Filterable (TSS)	SM 2540 D Gravimetric, 105° C	5	
Volatile (TVS)	EPA M160.4 Gravimetric, 550° C	10	
Resistivity (requires conductivity)	SM 2510 B (Meter)	1 (umhos-cm)	
Specific Gravity	SM 2710 F	0.001	
Turbidity	EPA M180.1 Nephelometric	0.1 (NTU)	



**Water and Wastewater
Inorganic
Sample Preparation**

Parameter	Method	Price/Sample
Lab Filtration *	SM 3030 B	
Lab Filtration & Acidification *	SM 3030 B	
Lab Filtration (Radiochemistry)*	SM 3030 B	

***(Difficult to filter, high TDS samples may require more then one filtration and additional charge)**

Metals Digestions - Hot Plate/Block Technique

Total Metals (ICP & ICP-MS)	EPA M200.2
Total Metals - (ICP & ICP-MS)	EPA M3010-A
Total Recoverable Metals-(ICP & ICP-MS)	EPA M3005-A
Total Recoverable Metals-(ICP & ICP-MS)	EPA M200.2
Potentially Dissolved Metals-(ICP & ICP-MS)	Colorado 5 CCR 1002-31.5.31 (2009)

Metals Analysis

Parameter	Method	MDL (ug/L)	Price/Sample
ICP-MS			
One-time per sample per species setup charge (if < 4 parameters requested)			
Routine parameters listed on page 7	EPA M200.8/6020		see page 7
Rare Earth Elements ¹	EPA M200.8/6020		
Cerium (Ce), Dysprosium (Dy), Erbium (Er), Europium (Eu), Gadolinium (Gd), Holmium (Ho)			
Lanthanum (La), Lutetium (Lu), Neodymium (Nd), Praseodymium (Pr), Scandium (Sc)			
Samarium (Sm), Terbium (Tb), Thulium (Tm), Yttrium (Y) & Ytterbium (Yb)		0.1	
ICP			
One-time per sample per species setup charge (if < 4 parameters requested)			
Routine parameters listed on page 7	EPA M200.7/6010B		see page 7
Mercury			
CVAA Mercury (Hg)	EPA 245.1 or 7470A	0.2	
CVAFS Mercury (Hg) ²	EPA 1631E	0.0002	
ICP-MS Mercury (Hg)	EPA M200.8/6020	0.025	
Hydride Generation AA (HGAA)			
HGAA Selenium (Se)	SM 3114 B & C	1	
HGAA Selenium Speciation	SM 3114 B & C		
Includes +6, +4, dissolved and organic		1	
Arsenic			
HPLC Arsenic Speciation	HPLC-ICP-MS (Includes As III & As V)	0.5	

1. Includes Lanthanides, Scandium and Yttrium

2. 1 Field transfer blank is required for each project of 1-10 samples. Field transfer blanks are considered billable samples.



**Water and Wastewater
Inorganic**

Metals Detection Limits¹ in mg/L²

Instrument EPA Method	ICP 200.7/6010	ICP/MS ^{3,4} 200.8/6020	Other
Aluminum (Al)	0.03	0.005	-
Antimony (Sb)	0.03	0.0004	-
Arsenic (As)	0.04	0.0002	-
Barium (Ba)	0.003	0.0005	-
Beryllium (Be)	0.01	0.00008	-
Bismuth (Bi)	0.04	-	-
Boron (B)	0.01	0.001	-
Cadmium (Cd)	0.005	0.0001	-
Calcium (Ca)	0.1	-	-
Cesium (Cs)	-	0.0002	-
Chromium (Cr)	0.01	0.0005	-
Cobalt (Co)	0.01	0.00005	-
Copper (Cu)	0.01	0.0008	-
Gallium (Ga)	0.1	-	-
Iron (Fe)	0.02	0.005	-
Lead (Pb)	0.03	0.0001	-
Lithium (Li)	0.008	-	-
Magnesium (Mg)	0.2	-	-
Manganese (Mn)	0.005	0.0005	-
Mercury (Hg)	-	0.04 ug/L	0.0002 (CVAA)
Mercury (Hg) (Ultra Low Level) ⁵	-	-	0.2 ng/L (1631)
Molybdenum (Mo)	0.02	0.0002	-
Nickel (Ni)	0.008	0.0004	-
Potassium (K)	0.2	-	-
Phosphorous (P)	0.1	-	-
Scandium (Sc)	0.1	0.0001	-
Selenium (Se)	0.05	0.0001	0.001 (Hydride)
Silicon (Si)	0.1	-	-
Silica (SiO ₂)	0.214	-	-
Silver (Ag)	0.01	0.0001	-
Sodium (Na)	0.2	-	-
Strontium (Sr)	0.005	-	-
Sulfur (S)	0.25	-	-
Tellurium (Te)	-	0.001	-
Thallium (Tl)	0.1	0.0001	-
Thorium (Th)	-	0.001	-
Tin (Sn)	0.04	0.0008	-
Titanium (Ti)	0.005	-	-
Uranium (U)	-	0.0001	-
Vanadium (V)	0.005	0.0005	-
Zinc (Zn)	0.01	0.004	-

1. Any required dilution due to matrix or high TDS will increase the MDLs.
2. Unless otherwise noted.
3. MDLs for elements analyzed by ICP-MS are for a dissolved basis only.
4. See previous page for Rare Earth Elements (Ce, Dy, Er, Eu, Gd, Ho, La, Lu, Nd, Pr, Sc, Sm, Tb, Tm, Y and Yb)
5. 1 Field transfer blank is required for each project of 1-10 samples. Field transfer blanks are considered billable samples.



**Water and Wastewater
Organic**

Gas Chromatography Techniques

Parameter	Method	MDL (mg/L)	Price/Sample
TPH (Extractable), DRO	EPA M8015D/API-DRO GC/FID	0.1	
TPH (C6-C35)	TNRCC Method 1005	3-10	
TVH (Volatile), GRO, (TVPH)	EPA M8015D/API-GRO GC/FID	0.05	
BTEX	EPA M8021B GC/PID	0.2 - 0.4 (ug/L)	
BTEX/MTBE	EPA M8021B GC/PID	0.2 - 0.4 (ug/L)	
BTEX/TVH, GRO	EPA M8021B/8015D GC/PID/FID	0.001-0.05	
BTEX/MTBE/TVH, GRO	EPA M8021B/8015D GC/PID/FID	0.001-0.05	
Dissolved Gas (Methane Only)	RSK 175 GC/FID	0.009	
Dissolved Gases	RSK 175 GC/FID		
Methane, Ethane, Ethene, Butane and Propane		0.009-0.032	
Organochlorine Pesticides	EPA 8081A GC/ECD	0.01	
PCBs	EPA M8082 GC/ECD	0.001	
Chlorophenoxy Acid Herbicides	EPA M615/8151 GC/ECD		

Gas Chromatography/Mass Spectrometry Techniques

Parameter	Method	MDL (ug/L)	Price/Sample
Volatiles, Low-Level (VOC)	EPA M624/8260B GC/MS	0.2 - 1	
Semi-Volatiles (SVOC)(BNAs)	EPA M625/8270C GC/MS	2 - 20	
SVOC Base/Neutral Extractables	EPA M625/8270C GC/MS	2 - 20	
SVOC Acid Extractables	EPA M625/8270C GC/MS	2 - 20	
PAHs	EPA M8270C GC/MS	2	
GC/MS Library Search (10 TICs/Fraction)	EPA M624/8260B GC/MS	N/A	
GC/MS Library Search (10 TICs/Fraction)	EPA M625/8270C GC/MS	N/A	
BTEX	EPA M8260B GC/MS	0.2 - 0.4	
Nonylphenols	ASTM D7065-06		
Monoethyloxalate, Diethyloxalate, Ocyphenol and Bisphenol A (BPA)		1-20 (mg/L)	

High Pressure Liquid Chromatography Techniques

Parameter	Method	MDL (ug/L)	Price/Sample
PAHs	EPA 8310, HPLC		



**Water and Wastewater
Radiochemistry**

Parameter	Method	LLD (pCi/L)	Price/Sample
Gamma Scan	EPA 901.1	various	
Gross Alpha/Beta	EPA 900.0/9310	2 & 4	
Gross Alpha/Beta (High TDS Samples)	EPA 600/00-02	3	
Lead - 210	Eichrom OTW01	4*	
Polonium - 210	HASL 300 PO-01-RC	1.5	
Radium - 226 (alpha emitting)	EPA 903.0 / 9315	1	
Radium - 226 (Drinking Water)	EPA 903.1	0.4	
Radium - 228 (Drinking Water)	EPA 904.0	1.5	
Radium - 228	EPA 9320	1.5	
Radon - 222	SM 7500 RN	25	
Thorium - Isotopic (228, 230, 232)	ESM 4108/4506 TOPO Extraction	0.3 - 0.6	
Tritium	EPA 906.0	1700	
Uranium, Isotopic (234, 235, 238)	Eichrom ACWO3	1	

Metals Analysis by ICP-MS

Parameter	Method	MDL (mg/L)	Price/Sample
Thorium, Total - 232	EPA 200.8/6020	1	
Uranium, Total - 238	EPA 200.8/6020	0.1	

1. An additional setup charge for ICP-MS analysis may apply. See page 6 for sample prep and setup charges.

*Lower MDLs may be achievable for certain radchem parameters with larger sample volume. Please contact an ACZ rep for more information

Please note, if ACZ is requested to lab filter for dissolved radiochemical parameters additional filtration charges may apply if the sample contains high TDS and is hard to filter or requires pressure filtration.



**Soil and Overburden
Inorganics
Sample Preparation**

Parameter	Method	Price/Sample
Air Dry	USDA No. 1, 1972	
Coarse Fragments (> 2.0mm)	ASA No. 9 15-4.2.2	
Crush / Pulverize (per 5 lb. sample)	USDA No. 1, 1972	
Crush / Pulverize (Ring & Puck)	EPA-600/2-78-054 3.1.3	
Crush / Pulverize (Ring & Puck) Low-Level	EPA-600/2-78-054 3.1.3	
Percent Moisture / Percent Solids	ASTM D2216-80/EPA M209F	
Screen / Pulverize	USDA No. 1, 1972	
Screen (per sieve fraction)	ASA No. 9 15-4.2.2	

Extractions

Parameter	Method	Price/Sample
AB-DTPA	ASA No. 9 3-5.2.3	
Acetic Acid	C.S.U., Soltanpour	
Acid Volatile Sulfide in Sediment (AVS)	EPA 08/91 Draft	
AVS with Simultaneously Extracted Metals	EPA 04/91 (14) Draft	
Bray P-1 Extraction	Soil Survey Method #6S3	
NH ₄ Acetate (Extractable)	USDA No. 60 (18)	
KCL	ASA No. 9 33-3.2.2	
Hot Water	ASA No.9 80-3	
Saturated Paste (includes Saturation %)	USDA No. 60 (2)	
Water	ASA No. 9 10-2.3.2	
TCLP Metals Extraction	EPA M1311	
TCLP SVOC Organics Extraction	EPA M1311	
Pests, Herbs and/or Semi-Volatiles		
TCLP VOC Organics Extraction (ZHE)	EPA M1311	
CA Waste Extraction Test (WET-STLC)	California Method, Title 22	
Meteoric Water Mobility Procedure	NDEP - MWMT (5/96) (Column)	
Meteoric Water Mobility Procedure	NDEP - MWMT (1990) (Bottle Roll)	
Column/Serial Batch Leach Studies	per client specifications	
Sequential Leach Studies	per client specifications	
SPLP	EPA M1312	
DI Leaching Procedure	EPA M1312 DI	
Humidity Cell Test	ASTM D5744-13	
Set-up		
Weekly Maintenance		
In Vitro Bioaccessibility (IVBA)	EPA M1340/9200.2-86	
As, Pb and other metals		



**Soil and Overburden
Inorganic
Digestions**

Parameter	Method	Price/Sample
EPA Method 3050B (Hot Block)	EPA M3050B	
EPA Method 3051 (Microwave)	EPA M3051A	
EPA Method 3051 (Waste Oil Digestion)	EPA M3051A	
EPA Method 3060 (For Hex-Chrome)	EPA M3060A	
USGS HCL / H ₂ O ₂	USGS Sediment	
Hydrofluoric/Nitric Acid Digestion	EPA M3052	

Chemical Properties

Parameter	Method	MDL (mg/Kg)	Price/Sample
Acid-Base Accounting (with Total Sulfur)	EPA 600/2-78-54		
Requires Neutralization Potential & Total Sulfur		1 (T CaCO ₃ /1000 T)	
Acid-Base Accounting (with Sulfur Forms)	EPA 600/2-78-54		
Requires Neutralization Potential & Sulfur Forms		1 (T CaCO ₃ /1000 T)	
Carbon			
Organic (Low Temperature Ignition)	EPA 600/2-78-54 3.2.14	0.1 (%)	
Total (TC)	ASA No. 9 29-2,2,4 IC Combustion (Ieco)	0.1 (%)	
Total Carbonate as CaCO ₃ (TIC)	ASA No. 9 29-2,2,4 IC Combustion (Ieco)		
(Requires TC and TOC)	Calculation	0.1 (%)	
Total Organic (TOC)	ASA No. 9 29-2,2,4 IC Combustion (Ieco)	0.1 (%)	
Cations, Soluble	EPA 6010B ICP		
Includes Ca, Mg, Na and K		various	
Chromium, Hexavalent	EPA M3060A/EPA 7196A	1	
Cation Exchange Capacity (CEC)	USDA No. 60 (19)	0.3 (meq/L)	
Chloride	SM4500Cl-E	5	
Conductivity	SM 2510 B/USDA 60 M2,21A,27A	0.01 (mmhos/cm)	
Cyanide			
Total	EPA M9012B	0.005	
Weak Acid Dissociable (WAD)	SM 4500 CN-I	0.005	
Exchangeable Cations (Ca, Mg, K & Na)	Calculation		
Requires: Cations via NH ₄ & Saturated Paste Extractions & Saturation %		N/A	
Exchangeable Sodium Percentage	USDA No. 60 (20 b)		
Requires: CEC, Na via NH ₄ & Saturated Paste Extractions & Saturation %		N/A	
Fluoride, Soluble	SM 4500 F-C/ASTM D 3761	0.1	
Halogens, Total Organic (TOX) Oil Matrix	D808 (05.01)	1	
Lime, Estimated (FIZZ) ³	EPA 600/2-78-054 3.2.3	N/A	
Lime Requirement (SMP Buffer)	EPA 600/2-78-054 3.2.1	0.5 (tons/hectare)	
Net Acid Generation Procedure (NAG)			
Single NAG	EGI 2002	0.1 to 1 Kg H ₂ SO ₄ /t	
Sequential NAG	EGI 2002	0.1 to 1 Kg H ₂ SO ₄ /t	
Australian NAG	Miller & Donahue (1997 ICARD)	0.1 to 1 Kg H ₂ SO ₄ /t	
NAG pH (Only)	EGI 2002	0.1 (units)	

1. Requires saturated paste extraction, see page 10.

2. Requires water extraction, see page 10.

3. Reported as text - N, SL, MO or ST



**Soil and Overburden
Inorganic
Chemical Properties**

Parameter	Method	MDL (mg/Kg)	Price/Sample
Neutralization Potential	EPA 600/2-78-054 3.2.3	0.1 (%)	
Neutralization Potential (Modified-no heat)	EPA 600/2-78-054 3.2.3 (MOD)	0.1 (%)	
Neutralization Potential Siderite Correction	EPA 600/2-78-054 3.2.3	0.1 (%)	
Nitrogen			
Ammonia	EPA M350.1	1	
Extractable NO ₃ /NO ₂	EPA M353.2	0.2	
Organic (Requires Total and Ammonia)	Calculation	0.1 (%)	
Soluble NO ₃ /NO ₂	EPA M353.2	0.02	
Total Kjeldahl	EPA M351.2	0.1 (%)	
pH	USDA No. 60 (21a)	0.1 (units)	
pH, Corrosivity	EPA M9045D/M9040C	0.1 (units)	
Phenols, Total	EPA M420.4	0.003	
Phosphorus, Extractable	EPA M365.1	0.02	
Phosphorus, Total	EPA M365.1	0.2 (%)	
Sodium Adsorption Ratio (SAR)	USDA No. 60 Calculation		
Requires Saturated Paste Extraction, Ca, Mg and Na		N/A (meq./ L ratio)	
Sulfur			
Forms (Coal/Overburden)	EPA 600/2-78-054 3.2.4 (Sobek)		
Includes Organic, Pyritic, Sulfate and Total		0.01 (%)	
Forms (Hard Rock w/heat)	EPA 600/2-78-054 3.2.4 (MOD Sobek)		
Includes Organic, Pyritic, Sulfate and Total		0.01 (%)	
Forms (Nevada Modified Sobek)	EPA 600/2-78-054 3.2.4 & 3.2.6		
Includes Organic, Pyritic, Sulfate, Hot Water Soluble Sulfate and Total		0.01 (%)	
Total	ASTM D-4239-85C (LECO)	0.01 (%)	
Sodium Carbonate Insoluble	ASTM D-4239-85C, D1915	0.01 (%)	
Water Insoluble	ASTM D-4239-97C (LECO)	0.01 (%)	
Sulfate, Soluble	SM 4500 SO4-D	10	

1. Requires KCL extraction, see page 10.
2. Requires water extraction, see page 10.
3. Requires AB-DTPA extraction, see page 10.



Soil and Overburden
Physical Properties

Parameter	Method	MDL (%)	Price/Sample
Color	Munsel Color chart	N/A	
Organic Matter	USDA No. 60 (24)	0.3	
Organic Matter - Ignition	EPA M600/2-78-054 3.2.14	0.3	
Particle Size			
Dry Sieve (per fraction)	ASA No. 9 15-4.2.2	0.1	
Wet Sieve (per fraction)	ASA No. 9 15-4.2.2	0.1	
Hydrometer (< 2.0mm)	ASA No. 9 15-5		
Sand, Silt and Clay by percentage plus Texture Classification		0.1	
Rock Fragments (Cobble, Gravel, etc.)	ASA No. 9 15-4.2.2	0.1	
Very Fine Sand (Wet Sieve)	ASA No. 9 15-4.2.2	0.1	
Bulk Density	ASA No. 9 13-4 Clod Method	N/A	
Solids, Volatile (550°C)	SM 2540G	0.1	
Solids, Total	SM 2540B	0.1	

Inorganic
Metals Analysis

Parameter	Method	MDL (mg/Kg)	Price/Sample
ICP/MS			
Onetime per sample per species setup charge (if < 4 parameters are requested)			
Routine parameters listed on page 14	EPA M6020	see page 14	
Rare Earth Elements ¹	EPA M6020		
Cerium (Ce), Dysprosium (Dy), Erbium (Er), Europium (Eu), Gadolinium (Gd), Holmium (Ho)			
Lanthanum (La), Lutetium (Lu), Neodymium (Nd), Praseodymium (Pr), Scandium (Sc)			
Samarium (Sm), Terbium (Tb), Thulium (Tm), Yttrium (Y) & Ytterbium (Yb)		0.05	
ICP			
Onetime per sample per species setup charge (if < 4 parameters are requested)			
Routine parameters listed on page 14	EPA M6010B	see page 14	
Mercury Analysis			
CVAA Mercury (Hg)	EPA M7471A/7470A	0.02	
Direct Combustion AA	EPA M7473	2-5 (ng/g) ²	
ICP-MS Mercury (Hg)	EPA 6020	0.05 (ug/L)	
Hydride Generation AA (HGAA) ³			
HGAA Selenium (Se)	EPA M7742 MOD-AA	0.001 (mg/L)	

1. Includes Lanthanides, Scandium and Yttrium

2. Lower detection limit available

3. TCLP, SPLP or MWMP extraction required



**Soil and Overburden
Inorganic**

Metals Detection Limits¹ in mg/Kg²

Instrument EPA Method	ICP 200.7/6010	ICP/MS ³ 200.8/6020	Other
Aluminum (Al)	3	0.5	-
Antimony (Sb)	3	0.2	-
Arsenic (As)	4	0.1	-
Barium (Ba)	0.3	0.25	-
Beryllium (Be)	1	0.025	-
Bismuth (Bi)	4	-	-
Boron (B)	1	0.25	-
Cadmium (Cd)	0.5	0.05	-
Calcium (Ca)	10	-	-
Cesium (Cs)	-	0.1	-
Chromium (Cr)	1	0.25	-
Cobalt (Co)	1	0.025	-
Copper (Cu)	1	0.25	-
Gallium (Ga)	10	-	-
Iron (Fe)	2	-	-
Lead (Pb)	3	0.05	-
Lithium (Li)	0.8	-	-
Magnesium (Mg)	20	-	-
Manganese (Mn)	0.5	0.25	-
Mercury (Hg)	-	0.025	0.02 (CVAA)
Mercury (Hg) Direct Combustion	-	-	2-5 ng/g (EPA 7473)
Molybdenum (Mo)	2	0.25	-
Nickel (Ni)	0.8	0.3	-
Potassium (K)	20	-	-
Phosphorous (P)	10	-	-
Scandium (Sc)	10	0.05	-
Selenium (Se)	5	0.05	-
Silicon (Si)	10	-	-
Silica (SiO ₂)	21.4	-	-
Silver (Ag)	1	0.025	-
Sodium (Na)	20	-	-
Strontium (Sr)	0.5	-	-
Sulfur (S)	0.25	-	-
Tellurium (Te)	-	0.5	-
Thallium (Tl)	10	0.05	-
Thorium (Th)	-	0.5	-
Tin (Sn)	4	0.05	-
Titanium (Ti)	0.5	-	-
Uranium (U)	-	0.05	-
Vanadium (V)	0.5	0.1	-
Zinc (Zn)	1	1	-

1. Any required dilution due to matrix or high TDS will increase the MDLs.

2. Unless otherwise noted.

3. See previous page for Rare Earth Elements (Ce, Dy, Er, Eu, Gd, Ho, La, Lu, Nd, Pr, Sc, Sm, Tb, Tm, Y and Yb)



**Soil and Overburden
Organic**

Gas Chromatography Techniques

Parameter	Method	MDL (mg/Kg)	Price/Sample
TPH (Extractable), DRO	EPA M8015D/API-DRO GC/FID	3.33	
TVH (Volatile), GRO, (TVPH)	EPA M8015D/API-GRO GC/FID	0.05	
BTEX	EPA M8021B GC/PID	1-2 (ug/Kg)	
BTEX/MTBE	EPA M8021B GC/PID	0.2-0.4 (ug/Kg)	
BTEX/TVH, GRO	EPA M8021B/8015D GC/PID/FID	0.001-0.05	
BTEX/MTBE/TVH, GRO	EPA M8021B/8015D GC/PID/FID	0.001-0.05	
Organochlorine Pesticides	EPA M8081A GC/ECD	0.01	
PCBs	EPA M8082 GC/ECD	0.033	
Chlorophenoxy Acid Herbicides	EPA M8151 GC/ECD		

Gas Chromatography/Mass Spectrometry Techniques

Parameter	Method	MDL (ug/Kg)	Price/Sample
Volatiles (VOC)	EPA M8260B GC/MS	3 - 10	
Semi-Volatiles (SVOC)	EPA M8270C GC/MS	66.6-333.3	
SVOC Base/Neutral Extractables	EPA M8270C GC/MS	66.6-333.3	
SVOC Acid Extractables	EPA M8270C GC/MS	66.6-333.3	
PAHs	EPA M8270C GC/MS	66.6	
GC/MS Library Search (10 TICs/Fraction)	EPA M8260B GC/MS	N/A	
GC/MS Library Search (10 TICs/Fraction)	EPA M8270C GC/MS	N/A	
BTEX	EPA M8260B GC/MS	0.2 - 0.4	

High Pressure Liquid Chromatography Techniques

Parameter	Method	MDL (mg/Kg)	Price/Sample
PAHs	EPA 8310, HPLC		



**Soil and Overburden
Radiochemistry**

Parameter	Method	LLD (pCi/g)	Price/Sample
Gamma Scan	ESM 901.1	various	
Gross Alpha/Beta	EPA 9310 modified	1 & 2	
Lead - 210	Eichrom OTW01	10	
Polonium - 210	HASL Po-01-RC	1.5	
Radium - 226 (alpha emitting)	EPA M9315	2	
Radium - 226	EPA M903.1	0.5	
Radium - 228	EPA M9320 / 904.0	3	
Thorium - Isotopic (228, 230, 232)	ESM 4506 TOPO Extraction	0.3 - 0.6	
Uranium - Isotopic (234, 235, 238)	Eichrom ACW 3	0.4	

All Soils radchem pricing listed above does not include sample prep or digestion charges see page 11.

No digestion charge is needed for gamma.

Metals Analysis by ICP-MS

Parameter	Method	MDL (mg/Kg)	Price/Sample
Thorium, Total	EPA 200.8/6020	0.5	
Uranium, Total	EPA 200.8/6020	0.05	

1. An additional setup charge for ICP-MS analysis may apply. See page 13 prep and setup charges.

*Lower MDLs may be achievable for certain radchem parameters with larger sample volume. Please contact an ACZ rep for more information.



Biological Tissue Analysis
Inorganic and Organic
Sample Preparation

Parameter	Method	Price/Sample
<i>Flora</i>		
Dry	Air Dry at 40° C	
Grind/Pulverize	Willy Mill or Riesch Knife Mill	
Extraction	EPA 600/4-81-055	
Ash	USNRC Guidelines	
Moisture Percent	M209F, Gravimetric - 105 C	

Fauna

Dissection	EPA 600/4-81-055	
Dissection	USDA/CDC Level III	
	USDA/CDC Biosafety Level III, i.e. Deer Mice	
Extractions - Metals	EPA 600/4-81-055	
Extractions/Cleanup - Organics	EPA M5000 series	
Digestions	EPA M200.3/3051	

Analysis

Parameter	Method	MDL (mg/Kg)	Price/Sample
<i>Metals</i>			
ICP/MS	EPA M6020	Available on request	
ICP	EPA M6010B	Available on request	
CVAA	EPA M7470A	Available on request	
Direct Combustion AA	EPA M7473	Available on request	

Nutrients

N, P, K, SO ₄ , B, cations, etc.	EPA M300 series, ASA, etc.	Available on request	
---	----------------------------	----------------------	--

Organics

Organochlorine Pesticides	EPA M8081A	Available on request	
PCBs	EPA M8082	Available on request	

Trace Metals: Urinalysis, Bioassay and Whole Blood

Parameter	Method	MDL (mg/L)	Price/Sample
<i>Metals</i>			
ICP/MS	EPA M200.8	Available on request	



**Solid and Hazardous Waste - RCRA
Inorganic and Organic
Hazardous Waste Characteristics**

Parameter	Method	Aqueous	Non-Aqueous
Ignitability			
Liquid (Flash Point / Ignitability)	EPA M1010A		
Solid (Ignitability)	EPA M1030		
Corrosivity			
pH	EPA M9040C/9045D		
Toxicity Characteristic Leaching Procedure (TCLP) Extractions (filtration only if <0.5% solids)			
Volatile Organics	EPA M1311		
Zero Headspace Extraction			
Semi-Volatile Organics	EPA M1311		
SVOCs and/or Pesticides and/or Herbicides			
Metals	EPA M1311		
Toxicity Characteristic Leaching Procedure (TCLP) Analyses			
Volatile Organics (20 compounds)	EPA M8260B		
Metals (8 RCRA elements) ¹	EPA M6010B/7470A		
Semivolatile Organics (16 compounds)	EPA M8270C		
Pesticides (6 compounds)	EPA M8081A		
Miscellaneous			
Oil/Solid/Water Percentages	ASTM D-96 & 95		
Percent Water (Toluene Co-Distillation)	ASTM D-95		
BTU Content (Parr Bomb)	ASTM D-240(05.01)		
PCBs in Transformer oils	EPA M8082		
PCBs on Wipes	EPA M8082		
PCBs in Sludge	EPA M8082		
Paint Filter Test (PFLT)	EPA M9095		
Waste Oil Parameters			
Flash Point	EPA M1010		
Metals (As, Cd, Cr, Pb and digestion)	EPA M6010/7000		
Total Halogens (TOX)	ASTM D808		
PCBs	EPA M8082		

1. 8 TCLP Metals quoted by ICP and CVAA; ICP-MS available at an additional charge.



**Solid and Hazardous Waste - RCRA
Inorganic and Organic**

Underground Storage Tank (UST) Analyses

Parameter	Method	Aqueous	Non-Aqueous
Gasolines/Aviation Fuel/Condensates			
Gasoline Range Organics (GRO)	EPA M8015D GC/FID		
Total Volatile Hydrocarbons (TVH)	EPA M8015D GC/FID		
Total Volatile Petroleum Hydrocarbons	EPA M8015D GC/FID		
BTEX/MTBE	EPA M8021B GC/PID		
TVH/BTEX	EPA M8015D/M8021B		
Volatile Alcohols (Ethylene Glycol)	EPA M8015A-Mod		
Paint Filter Liquids Test (PFLT)	EPA M9095		
Diesels/Jet Fuels/Fuel Oils/Kerosene			
Diesel Range Organics (DRO)	EPA M8015D GC/FID		
Total Petroleum Hydrocarbons (TPH)	EPA M8015D GC/FID		
Paint Filter Liquids Test (PFLT)	EPA M9095		
Lubricating Oils/Bunker Oils/Waste Oils/Greases			
Oil & Grease	EPA M1664A		
Oil & Grease (Soxhlet Extraction)	EPA M9071B		
Diesel Range Organics (DRO)	EPA M8015D GC/FID		
PAHs	EPA M8310		
Paint Filter Liquids Test (PFLT)	EPA M9095		
PCBs in Waste Oils	EPA M8082A		

Appendix IX Constituents

Parameter	Method	Aqueous	Non-Aqueous
Metals (17) and Inorganics (2)			
Phenols			
Organochlorine Pesticides			
Organophosphorus Pesticides			
Chlorinated Herbicides			
Volatile Organics (with TIC)			
Semivolatile Organics (with TIC)			
Dioxins & Furans			



**Regulatory Parameters and Packages
Inorganic and Organic**

California Administrative Code (CAC)

Parameter	Method	Aqueous	Non-Aqueous
Waste Extraction Test (WET)	CAL Title 22		
Metals (18) (TTLC)	EPA M6010B/6020/7470A		
Fluoride	SM 4500 F-C		
Chlorinated Pesticides (13) / PCBs	EPA M8081A/8082		
Chlorophenoxy acid Herbicides (2)	Subcontractor specific		
Total Organic Lead	Subcontractor specific		
Dioxin (2,3,7,8-TCDD)	Subcontractor specific		
Trichloroethylene	EPA M8260B		
Pentachlorophenol	EPA M8270C		

Clean Water Act (CWA)

Parameter	Method	Aqueous	Non-Aqueous
Metals (13)	EPA M200 series		
Cyanide, Total	EPA M335.4 Auto Color		
Phenols, Total	EPA M420.4 Auto Color		
Volatile Organics (28)	EPA M624 GC/MS		
Semivolatile Organics (57)	EPA M625 GC/MS		
Organochlorine Pesticides/PCBs (25)	EPA M608 GC/ECD		
Total Toxic Organics (TTO) Includes VOC, SVOC, Pests/PCBs	EPA M624/625/608		

Colorado Department of Public Health and Environment (CDPHE)

Parameter	Method	Aqueous	Non-Aqueous
Leaking Storage Tank Investigations			
TVPH	EPA M8015B-Mod		
TEPH	EPA M8015-Mod		
Oil & Grease	EPA1664A / EPA 9071A		
BTEX	EPA M8021B		
Municipal Solid Waste Landfill - GW Detection Monitoring			
Cations (Ca, Mg, K and Na)	EPA M6010B		
Anions (Alk, Cl, NO ₃ , NO ₂ and SO ₄)	EPA M300 series, varies		
Metals (15 elements)	EPA M6010B/7000's		
Volatile Organic Compounds	EPA M8260B		



**Regulatory Parameters and Packages
Inorganic and Organic**

Nevada Division of Environmental Protection (NDEP)

Parameter	Method	Price/Sample
Profile I Metals		
Total Metal Digestion (water)	EPA M200.2	
ICP Metals	EPA M200.7	
Al, Ba, Be, Ca, Cr, Cu, Fe, Mg, Mn, Ni, K, Ag, Na and Zn		
ICP-MS Metals	EPA M200.8	
Sb, As, Cd, Pb, Se and Tl		
Mercury	EPA M 245.1 Auto CVAA	
Profile I Inorganics		
Alkalinity (Total CaCO ₃ and HCO ₃)	SM 2320 B Titrametric	
Chloride	SM 4500 Cl-E Auto Color	
Cyanide, WAD	SM 4500 CN-I	
Fluoride	SM 4500 F-C	
Nitrate/Nitrite	EPA M353.2 Auto Color	
Nitrogen, Total Kjeldahl (TKN)	EPA M351.2	
pH	SM 4500 H+B Meter	
Sulfate	D 516-02/07, Turbidimetric	
Total Dissolved Solids	SM 2540 C Gravimetric	
Lab Filtration and Acidification (x2)	(0.45 um filter)	
Profile II Metals		
Total Metal Digestion	EPA M200.2	
ICP Metals	EPA M200.7	
Al, B, Ba, Be, Bi, Ca, Cr, Co, Cu, Fe, Ga, Li, Mg, Mn, Mo, Ni, K, Sc, Ag, Na, Sr, Sn, Ti, V, Zn		
ICP-MS Metals	EPA M200.8	
Sb, As, Cd, Pb, Se and Tl		
Mercury	EPA M 245.1 Auto CVAA	
Profile II Inorganics		
Alkalinity (Total CaCO ₃ and HCO ₃)	SM 2320 B Titrametric	
Chloride	SM 4500 Cl-E Auto Color	
Cyanide, WAD	SM 4500 CN-I	
Fluoride	SM 4500 F-C	
Nitrate/Nitrite	EPA M353.2 Auto Color	
Nitrogen, Total Kjeldahl (TKN)	EPA M351.2	
pH	SM 4500 H+B Meter	
Phosphorus, Total (with Digestion)	EPA M365.1 Auto Color	
Sulfate	D 516-02/07, Turbidimetric	
Total Dissolved Solids	SM 2540 C Gravimetric	
Lab Filtration and Acidification (x2)	(0.45 um filter)	



**Regulatory Parameters and Packages
Inorganic and Organic**

Nevada Division of Environmental Protection (NDEP)

Parameter	Method	Price/Sample
Meteoritic Water Mobility Procedure (MWMP)		
Extraction	NDEP - MWMT (5/96) (Column) NDEP - MWMT (Bottle Roll)	
Total Metal Digestion	EPA M200.2	
ICP Metals	EPA M200.7	
Al, Ba, Be, Ca, Cr, Cu, Fe, Mg, Mn, Ni, K, Ag, Na and Zn		
ICP-MS Metals	EPA M200.8	
As, Cd, Pb, Se and Tl		
Mercury	EPA M245.1 CVAA	
Inorganics		
Alkalinity (Total CaCO ₃)	SM 2320 B Titrametric	
Chloride	SM 4500 Cl-E Auto Color	
Cyanide, WAD	SM 4500 CN-I	
Fluoride	SM 4500 F-C	
Nitrate(as N), Nitrite and Nitrate/Nitrite	EPA M353.2 Auto Color	
pH	SM 4500 H+B Meter	
Phosphorus	EPA 365.1 Auto Color	
Sulfate	D 516-02/07, Turbidimetric	
Total Dissolved Solids	SM 2540 C Gravimetric	
Acid Generation Potential (State of Nevada Modified Sobek)		
Static Testing		
Neutralization Potential Endpoint 8.3 (ANP)	EPA 600/2-78-054 3.2.3	
Neutralization Potential Siderite Correction	EPA 600/2-78-054 3.2.3	
Acidification Potential - Total Sulfur (AGP)	ASTM D-4239-85C (LECO)	
Acidification Potential - Sulfide Sulfur (AGP)	EPA 600/2-78-054 3.2.3	
Includes Organic, Pyritic, Sulfate, Hot Water Soluble Sulfate and Total		
Custom Static Testing is available	Call ACZ	
Kinetic Testing		
Shake Flasks	per client specifications	
Soxhlet Reactor	per client specifications	
Humidity Cell Test (Set-up and Maintenance)	per client specifications	
Humidity Cell Test (Prep and Analysis)	per client specifications	
Column Leach	per client specifications	

1. Sample preparation may be necessary, see page 10.



**Regulatory Parameters and Packages
Inorganic (IOC's)**

Safe Drinking Water Act (SDWA); National Primary Drinking Water Standards

Parameter	Method	MCL (mg/L)	Price/Sample
Primary Drinking Water Regs (PDWR)			
Arsenic	EPA M200.8	0.05	
Fluoride Rule			
Fluoride	SM 4500 F-C	4	
Surface Water Treatment Rule			
Turbidity	EPA M180.1	TT	
Phase II & IIB			
Asbestos (>10um)	TEM	7 (MFL)	
Barium	EPA M200.7	2	
Cadmium	EPA M200.8	0.005	
Chromium (total)	EPA M200.7	0.1	
Mercury	EPA M245.1	0.002	
Nitrate/Nitrite (as N)	EPA M353.2	N/A	
Nitrite (as N)	EPA M353.2	1	
Nitrate (as N)	EPA M353.2 Calculation		
Requires NO ₂ /NO ₃ and NO ₂		10	
Selenium	EPA M200.8	0.05	
Lead and Copper Rule			
Copper	EPA M200.8	1.3	
Lead	EPA M200.8	0.015	
Phase V			
Antimony	EPA M200.8	0.006	
Beryllium	EPA M200.8	0.004	
Cyanide	EPA M335.4	0.2	
Nickel	EPA M200.8	0.0002	
Thallium	EPA M200.8	0.002	

1. An additional setup charge for ICP & ICP-MS analysis may apply.



**Regulatory Parameters and Packages
Inorganic (IOC's)**

Safe Drinking Water Act (SDWA); National Primary Drinking Water Standards

Parameter	Method	MCL (mg/L)	Price/Sample
Unregulated			
Sulfate	SM 4500 SO4-D, Gravimetric	400/500	
Aluminum	EPA M200.7	0.05 - 0.2	
Chloride	SM 4500 CL-E	250	
Color	EPA M110.2	15 (cu)	
Copper	EPA M200.8	1	
Corrosivity (Langlier Index)	SM 2330B Calculation		
Requires Alkalinity, Ca, pH, TDS and Field Temperature		N/A	
Fluoride	SM 4500 F-C	2	
Iron	EPA M200.7	0.3	
Manganese	EPA M200.8	0.05	
pH	SM 4500 H+B meter	6.5 - 8.5	
Silver	EPA M200.8	0.1	
TDS	SM 2540 C	500	
Zinc	EPA M200.8	0.002	

Radiological Standards (Interim and Proposed)

Alpha emitter	EPA M900.0	15	
Beta/photon emitters	EPA M900.0	4 (mrem/yr)	
Radium 226	EPA M903.1	5	
Radium 228	EPA M904.0	5	
Radon	SM 7500 RN	200	
Uranium by ICP-MS	EPA M200.8		
Includes digestion and ICP-MS setup charge		0.03 (mg/L)	

1. An additional setup charge for ICP & ICP-MS analysis may apply.



Administrative Services

Data Reporting Formats

	Standard Report	QC Report	Extended Report
Cover Page	X	X	X
Case Narrative (if required)	X	X	X
Analytical Results Report			
Parameter Name	X	X	X
EPA Method	X	X	X
Analytical Results	X	X	X
Qualifiers	X	X	X
Units	X	X	X
MDL / PQL	X	X	X
Analysis Date	X	X	X
Analyst	X	X	X
Reference Table	X	X	X
Sample Receipt Form	X	X	X
Chain of Custody	X	X	X
Attachments (if required)	X	X	X
QC Summary			
Initial Calibration Summary		X	X
Daily Calibration Check Results			X
Blank Results		X	X
Spike Results		X	X
LCS Results		X	X
Analytical/Instrumentation Logs			X
Raw Data			X
Calibration Chromatograms			X

Data Reporting Charges

Available Formats	Rates
Standard Report	N/C
QC Report (ACZ Standard QC)	N/C
Extended Report	By Quote
Full CLP-like Level IV reporting	By Quote
Custom reports per client specifications	By Quote
Hard Copy by Mail	By Quote



Administrative Services

Electronic Data Deliverables

Available Formats	Rates
ACZ Default (MS Excel spreadsheet or MS Access database table)	
Custom EDD formats per client specifications	

Labor

Description	Rates
Expert Witness, plus travel	
Officer	
Section Supervisor	
Project Manager	
Senior Chemist	
Chemist	
Technician	
Clerical	

Rush Turnaround

Description	Rates
1 Working Day Turnaround	
2 Working Days Turnaround	
3 Working Days Turnaround	
4 Working Days Turnaround	
5 Working Days Turnaround	
7 Working Days Turnaround	
10 Working Days Turnaround	
15 Working Days Turnaround	

Sample Storage, Return and Disposal

Description	Rates
Long Term Storage - Non-Hazardous Waste Samples	
Disposal - Hazardous Waste Samples	
Disposal - Non-Hazardous Waste Samples	
Return/Ship Samples Back to Client	

Small Sampling Event

ACZ reserves the right to add a surcharge for any project with less than five samples. Minimum charge is \$150.00
Please contact your customer service representative or project manager for details.



Terms & Conditions

General Agreement

These Terms and Conditions, together with any additions or revisions which may be agreed to in writing by ACZ, embody the whole agreement of the parties in the absence of a signed and executed contract between ACZ Laboratories, Inc., hereinafter referred to as ACZ, and Client. They shall supersede all previous communications, representations, or agreements, either verbal or written, between the parties. ACZ specifically rejects all additional, inconsistent or conflicting terms, whether printed or otherwise set forth in any purchase order or other communication from the Client to ACZ. The invalidity or unenforceability, in whole or in part of any provision, term or condition hereof shall not affect in any way the validity or enforceability of the remainder of the Terms and Conditions, the intent of the parties being that the provision is severable. No waiver by ACZ of any provision, term or condition hereof or of any breach by or obligation of Client hereunder shall constitute a waiver of such provision, term or condition on any other occasion or a waiver of any other breach by or obligation of Client, unless specifically set forth in writing and executed on behalf of ACZ by a duly authorized officer. This agreement shall be administered and interpreted under the laws of the State of Colorado.

Order of Service

Part A

- A1 ACZ will accept orders by submission of written purchase order, by telephone or by contractual agreement. Telephone orders should be subsequently confirmed in writing to avoid potential misunderstandings. Submission of a telephone or written purchase order constitutes Client's acceptance of these general terms and conditions.

- A2 ACZ will provide Client with all necessary sample containers, coolers, labels and chain of custody documents at no additional charge provided these supplies are used for samples to be analyzed by ACZ. Sample containers provided by ACZ will include appropriate preservatives.

- A3 ACZ will ship supplies to Client's US destination at ACZ expense using UPS ground services. Expedited services will be invoiced at cost to Client. Client pays for return shipping unless otherwise agreed to in a signed contract between ACZ and Client. Expense of international shipping of sample containers will be by a contract basis only between ACZ and Client.

- A4 All orders are subject to a minimum charge of \$125.00.



Terms & Conditions

Credit Application & Payment Terms

Part B

- B1 Payment terms are Net 30 Days from invoice date unless otherwise executed by written contract. A service charge of 1.5% (18% per annum) will be made on all unpaid invoices 31 (or more) days old. Should it be necessary to assign account balance to a collection agency or attorney for legal action, all subsequent charges and legal fees shall be paid by Client. Should litigation become necessary, Client agrees that the venue shall be Routt County, State of Colorado. ACZ reserves the right to require payment prior to release of data. If payment terms are exceeded, ACZ has no obligation, and will not defend, reproduce, return or supplement data results.

- B2 Services performed by ACZ will be in accordance with quoted prices or as stated in the fee schedule, which is subject to change without notice. It is preferred that the client calls an ACZ representative prior to submitting samples to verify price and turnaround time.

- B3 In applying for credit, the Client certifies that the information (along with any other information submitted) is true and correct, and will update information that materially changes. All information received shall be treated confidentially by ACZ, and will be used only for the purpose of establishing the amount and conditions of extending commercial credit (not for consumer purposes). Client hereby authorizes all bank and trade references listed in this account application and agreement to release all information, verbal or written, to ACZ and to allow ACZ to utilize any other sources of credit information which ACZ deems reliable. Subsequent credit inquires may be completed by ACZ in connection with any update, renewal or extension of credit. ACZ reserves the right to suspend or terminate credit at any time.

- B4 A consulting firm who authorizes ACZ to perform work for Client must be responsible to have Client submit a signed credit application to ACZ that ACZ approves in advance of performing the work. If Client credit is approved, said Client shall elect to either a) be billed directly or, b) have work billed to consulting firm. Regardless if the Client or the consultant is billed, Client accepts final responsibility for payment to ACZ.

Sample Receipt, Delivery of Services and Sample Storage

Part C

- C1 Prior to receipt of samples at ACZ, Client is responsible for the entire risk of loss or any damage to samples. In no event will ACZ assume any responsibility or liability for the action or inaction of any carrier shipping or delivering samples to or from ACZ's premises.

- C2 ACZ reserves the right at any time to refuse delivery of, or to revoke acceptance of, any samples which, in the judgment of ACZ, represents a health, safety or environmental risk due to handling, transporting or processing of such samples.

- C3 All NRC licensed material and/or samples not meeting ACZ's 250 uRad/Hr screening levels, will be returned to the facility of origin at Client's expense. Any licensed material sent to ACZ for analytical testing must be identified on the COC or Client may incur additional costs including any labor and return fees.



Terms & Conditions

Sample Receipt, Delivery of Services and Sample Storage

Part C

- C4 ACZ utilizes analytical methodologies appropriate for sample matrices and accepted by EPA, USGS, ASTM, ASA, Standard Methods and other professional associations. Unless specifically mandated by Client, ACZ reserves the right to use alternative, equivalent methods. It is the responsibility of Client to verify with the regulatory agency that ACZ's quoted methods are acceptable for their statement of work.
- C5 Upon receipt of samples, ACZ will use its best efforts to comply with storage, processing and analytical holding time limits as stated in applicable EPA, state guidelines, as requested by Client. However, unless specifically made part of a written agreement between ACZ and Client, such time limits and analytical turnaround times are not guaranteed.
- C6 In accordance with ACZ sample acceptance policy, a client must use ACZ's Chain of Custody. ACZ's Chain of Custody contains specific information regarding the criteria that determines whether or not samples will be accepted for analysis. If an alternative Chain of Custody is used, ACZ will make the best effort possible to accept and analyze samples; however, ACZ will not be held liable for any issues that may arise from the use and submission of an improper Chain of Custody form. Likewise, ACZ will not be held liable when an ACZ Chain of Custody is not filled out completely or properly. It is the responsibility of the Client to ensure that all personnel conducting sampling activities are made aware of the requirements stated within Section C6 of ACZ's Terms & Conditions.
- C7 ACZ reserves the right to be compensated for instrument down time or damage caused by highly contaminated samples sent to ACZ for trace level analysis (PPM, PPB or PPT) due to percent level or pure product being shipped to ACZ without warning the lab on the COC or in writing to a Project Manager.
- C8 Preliminary results, verbal or written, may be given to Client in advance of the final written report. Preliminary data has not been fully reviewed in accordance with ACZ's Quality System and is subject to change without notice.
- C9 ACZ will at times subcontract analytical services ordered by Client. In all cases, Client will be informed prior to subcontracting out the work to another laboratory. ACZ assumes no liability for any subcontracted services except as specifically provided for in Part D.
- C10 ACZ will store all samples 30 days from date of invoice. At the end of this period, samples will be disposed of properly in compliance with applicable laws or returned to Client. Client will be responsible for shipping and handling charges associated with the return shipping of samples. Client will be billed a surcharge for sample disposal as agreed upon in any written agreement or at the rate quoted in the Administrative Services section of the current ACZ fee schedule. Any special storage, disposal or return requirements of Client must be outlined on the Chain of Custody prior to delivery of samples. Extended storage at Client's request will be invoiced as per the Administrative Services section of ACZ's fee schedule.
- C11 Client will be charged \$11.00 per sample disposal fee for all samples deemed to be hazardous.



Terms & Conditions

Warranties, Liabilities and Indemnification

Part D

- D1 ACZ warrants only that its services will fulfill obligations set forth in Sections C4 and C5 above. This warranty is the sole and exclusive warranty given by ACZ in connection with any services performed by ACZ or any results generated from such services. ACZ gives and makes no other representation or warranty of any kind, expressed or implied. Unless a specific contract or Statement of Work (SOW) is accepted by ACZ, no representative of ACZ is authorized to give or make any other representation or warranty or modify this warranty in any way.

- D2 The liabilities and obligations of ACZ and the remedies of Client in connection with any services performed by ACZ, will be limited to repeating the service performed or, at the sole option of ACZ, refunding in full or in part, fees paid by Client for such services. ACZ's obligation to repeat services will be contingent on the client providing, at the request of ACZ and at the client's expense, any additional samples necessary. Any reanalysis generating results consistent with the original results will be at Client's expense. Except as otherwise specifically provided herein, ACZ shall have no liability, obligation or responsibility of any kind for any losses, costs, expenses or other damages for any representation or warranty of any kind with respect to ACZ's service or results.

- D3 In no event shall ACZ have any responsibility or liability to Client for any failure or delay in performance by ACZ which results, directly or indirectly, from any cause or circumstances beyond the reasonable control of ACZ. Such causes or circumstances shall included, but are not limited to, acts of God, acts of Client, acts or order of any governmental authority, labor disputes, natural disasters, accidents, wars, difficulties or delays in transportation, mail or delivery services, inability to obtain from ACZ's usual sources sufficient services or supplies, or any other cause beyond ACZ's reasonable control.

- D4 All results provided by ACZ are strictly for the use of Client. ACZ is in no way responsible for use of such results by Client or third parties. All results should be considered in their entirety and ACZ is in no way responsible for the separation, detachment, or other use of any portion of the results.

- D5 Client represents and warrants that any sample delivered to ACZ will be preceded or accompanied by complete written disclosure of the presence of any hazardous substance known or suspected by Client. Client further warrants that any sample containing any hazardous substance that is delivered to ACZ will be packaged, labeled, transported and delivered properly and in accordance to applicable regulations.

- D6 The customer shall indemnify and hold harmless ACZ from and against any and all claims, suits, judgments, damages, losses, liabilities, expenses, payments, taxes, duties, fines and/or other costs (including but not limited to reasonable attorney's fees and liability to a third party) arising out of: a) the presence of hazardous substances in any sample of Client regardless of Client's compliance with paragraph D5 hereof, b) accidents occurring during the transport of any sample of Client, c) events or delays caused by Client or otherwise beyond ACZ's control, or d) negligence by Client in the use, evaluation, or application of results provided by ACZ.